
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002594
9:26 24/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:16
PRESSURIZING 2:22
STERILIZING 3:40
136.4 C 225kPa 3:40
136.3 C 224kPa 4:10
136.2 C 224kPa 4:40
136.4 C 224kPa 5:10
136.2 C 224kPa 5:40
136.2 C 225kPa 6:10
136.4 C 224kPa 6:40
136.4 C 226kPa 7:10
136.4 C 226kPa 7:10
VENTING 7:10
AIR DRYING 8:10
CYCLE COMPLETE 14:14

Digital Signature #
3D4F8227D620E5FF
