
Midmark M11 - v1.0.0
Total Cycles: 1680

Name: RITTER 1
Sterilizer ID: V2178198
02/28/2022 02:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.8 F 0.0
2:00 189.2 F 3.7
4:00 198.4 F 8.4
6:00 216.7 F 10.8
8:00 228.9 F 8.3
10:00 241.4 F 12.4
12:00 240.9 F 11.3
14:00 239.2 F 9.7
16:00 253.2 F 16.7
18:00 265.1 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.3 F 28.6
2:00 271.4 F 28.1
3:00 272.3 F 28.6
4:00 271.5 F 28.0

Min 271.4 F 27.9
Max 272.4 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.8 F 0.0
5:00 232.4 F 0.0
10:00 230.0 F 0.0
15:00 228.5 F 0.0
20:00 225.5 F 0.0
25:00 225.4 F 0.0
30:00 231.7 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 19:47
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:41
CYCLE COMPLETE

02/28/2022 03:38 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____