
Midmark M11 - v1.0.0
Total Cycles: 2648

Name: RITTER 1
Sterilizer ID: V2178198
12/06/2022 01:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.7 F 0.0
2:00 160.5 F 1.8
4:00 173.6 F 5.4
6:00 201.3 F 10.5
8:00 221.8 F 4.7
10:00 233.0 F 10.3
12:00 237.0 F 7.6
14:00 242.6 F 12.8
16:00 241.5 F 11.3
18:00 254.8 F 18.1
20:00 266.6 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.6 F 29.2
2:00 271.6 F 28.7
3:00 272.5 F 29.2
4:00 271.7 F 28.5

Min 271.5 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.3 F 0.0
5:00 202.7 F 0.0
10:00 216.8 F 0.0
15:00 221.4 F 0.0
20:00 224.0 F 0.0
25:00 226.3 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 21:32
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:59
CYCLE COMPLETE

12/06/2022 02:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____