

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5339

Name: RITTER 1  
Sterilizer ID: V2178198  
02/13/2025 10:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 153.8 F 0.0  
2:00 157.7 F 2.6  
4:00 168.9 F 6.2  
6:00 188.2 F 10.4  
8:00 201.7 F 3.9  
10:00 218.0 F 6.8  
12:00 231.2 F 11.5  
14:00 230.9 F 7.6  
16:00 243.8 F 13.4  
18:00 242.2 F 11.6  
20:00 254.8 F 18.1  
22:00 265.9 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.8 F 29.3  
2:00 271.8 F 28.5  
3:00 272.6 F 29.3  
4:00 272.4 F 28.9

Min 271.5 F 28.4  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.7 F 0.0  
5:00 196.5 F 0.0  
10:00 208.8 F 0.0  
15:00 212.9 F 0.0  
20:00 214.3 F 0.0  
25:00 218.2 F 0.0  
30:00 219.4 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 23:43  
STERILIZING: 4:00  
VENTING: 3:30  
DRYING: 30:00  
TOTAL CYCLE: 01:02:04  
CYCLE COMPLETE

02/13/2025 12:00 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_