

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5365

Name: RITTER 1  
Sterilizer ID: V2178198  
02/25/2025 04:28 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 140.8 F 0.0  
2:00 146.7 F 3.1  
4:00 173.4 F 7.3  
6:00 197.1 F 12.4  
8:00 211.9 F 5.1  
10:00 227.6 F 10.2  
12:00 231.7 F 7.2  
14:00 241.5 F 12.5  
16:00 240.2 F 10.8  
18:00 253.7 F 17.6  
20:00 265.7 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.5 F 29.1  
2:00 272.0 F 29.2  
3:00 272.2 F 28.8  
4:00 272.3 F 29.4

Min 271.5 F 28.5  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.8 F 0.0  
5:00 195.4 F 0.0  
10:00 202.7 F 0.0  
15:00 193.8 F 0.0  
20:00 205.6 F 0.0  
25:00 208.7 F 0.0  
30:00 213.7 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 21:39  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:00:04  
CYCLE COMPLETE

02/25/2025 05:28 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_