
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 000508
8:49 13/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:15
PRESSURIZING 3:17
STERILIZING 4:28
136.0 C 228kPa 4:28
136.1 C 230kPa 4:58
136.0 C 230kPa 5:28
136.3 C 231kPa 5:58
136.4 C 233kPa 6:28
136.4 C 232kPa 6:58
136.2 C 229kPa 7:28
136.4 C 232kPa 7:58
136.4 C 232kPa 7:58
VENTING 7:58
AIR DRYING 8:50
CYCLE COMPLETE 23:50

Digital Signature #
574C4B39469BA4C1
