
Midmark M11 - v1.0.0
Total Cycles: 3440

Name: RITTER 1
Sterilizer ID: V2178198
06/23/2023 04:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.7 F 0.0
2:00 181.8 F 3.2
4:00 202.7 F 7.7
6:00 224.1 F 13.0
8:00 230.4 F 8.6
10:00 239.4 F 11.0
12:00 243.0 F 12.2
14:00 241.6 F 11.2
16:00 255.1 F 18.1
18:00 266.5 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.5 F 28.8
2:00 271.4 F 28.5
3:00 272.6 F 28.9
4:00 271.5 F 28.3

Min 271.2 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.5 F 0.0
5:00 233.7 F 0.0
10:00 227.6 F 0.0
15:00 228.3 F 0.0
20:00 223.9 F 0.0
25:00 221.9 F 0.0
30:00 220.2 F 0.0
Drying Complete

FILLING: 2:17
HEATING: 19:33
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:59:12
CYCLE COMPLETE

06/23/2023 05:27 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____