

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1451

Name: RITTER 1  
Sterilizer ID: V2178198  
12/13/2021 01:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 211.2 F 0.0  
2:00 202.1 F 3.9  
4:00 201.1 F 9.3  
6:00 226.6 F 6.3  
8:00 232.9 F 11.9  
10:00 238.4 F 9.6  
12:00 235.7 F 8.4  
14:00 250.5 F 15.3  
16:00 263.0 F 22.5  
18:00 271.5 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.3  
1:00 272.1 F 28.6  
2:00 272.4 F 29.0  
3:00 271.4 F 28.2  
4:00 272.5 F 29.5  
  
Min 270.9 F 28.2  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.7 F 0.0  
5:00 229.2 F 0.0  
10:00 226.9 F 0.0  
15:00 229.9 F 0.0  
20:00 231.9 F 0.0  
25:00 228.6 F 0.0  
30:00 229.8 F 0.0  
Drying Complete

FILLING: 2:34  
HEATING: 18:06  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 30:00  
TOTAL CYCLE: 00:57:41  
CYCLE COMPLETE

12/13/2021 02:13 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_