

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2365

Name: RITTER 1  
Sterilizer ID: V2178198  
09/22/2022 05:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 207.6 F 0.0  
2:00 201.4 F 4.5  
4:00 200.9 F 9.4  
6:00 220.3 F 7.3  
8:00 230.8 F 9.9  
10:00 236.0 F 8.1  
12:00 244.1 F 13.3  
14:00 243.4 F 12.1  
16:00 256.6 F 19.1  
18:00 268.1 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.5  
1:00 272.4 F 29.3  
2:00 272.2 F 28.8  
3:00 271.3 F 28.5  
4:00 272.2 F 29.3  
  
Min 271.2 F 28.4  
Max 272.5 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.0 F 0.0  
5:00 227.9 F 0.0  
10:00 234.9 F 0.0  
15:00 231.3 F 0.0  
20:00 227.4 F 0.0  
25:00 227.4 F 0.0  
30:00 226.3 F 0.0  
Drying Complete

FILLING: 2:03  
HEATING: 19:19  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:37  
CYCLE COMPLETE

09/22/2022 06:41 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_