
Midmark M11 - v1.0.0
Total Cycles: 5408

Name: RITTER 1
Sterilizer ID: V2178198
03/11/2025 03:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 91.6 F 0.0
2:00 92.4 F 0.7
4:00 102.3 F 2.5
6:00 132.2 F 5.1
8:00 162.8 F 8.2
10:00 183.0 F 11.6
12:00 210.3 F 3.1
14:00 213.5 F 7.3
16:00 225.1 F 11.4
18:00 228.3 F 6.7
20:00 239.5 F 12.2
22:00 238.3 F 10.0
24:00 250.6 F 15.9
26:00 260.4 F 22.2
28:00 271.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 271.8 F 28.8
2:00 272.3 F 29.3
3:00 271.5 F 29.1
4:00 271.7 F 28.9

Min 271.1 F 28.7
Max 272.4 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.6 F 0.0
5:00 230.5 F 0.0
10:00 226.7 F 0.0
15:00 229.9 F 0.0
20:00 236.2 F 0.0
25:00 227.9 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 28:30
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:07:01
CYCLE COMPLETE

03/11/2025 04:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____