
Midmark M11 - v1.0.0
Total Cycles: 2644

Name: RITTER 1
Sterilizer ID: V2178198
12/06/2022 08:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 82.5 F 0.0
2:00 83.8 F 0.0
4:00 93.3 F 0.0
6:00 119.0 F 0.4
8:00 152.8 F 2.8
10:00 180.7 F 6.0
12:00 224.0 F 10.4
14:00 220.0 F 5.9
16:00 230.5 F 8.0
18:00 240.6 F 12.3
20:00 235.1 F 8.7
22:00 242.0 F 10.5
24:00 244.8 F 12.5
26:00 256.9 F 18.9
28:00 267.8 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.4 F 29.0
2:00 272.5 F 28.7
3:00 272.0 F 28.4
4:00 271.7 F 28.6

Min 271.6 F 28.4
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.5 F 0.0
5:00 211.2 F 0.0
10:00 198.8 F 0.0
15:00 198.1 F 0.0
20:00 196.0 F 0.0
25:00 189.6 F 0.0
30:00 184.5 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 29:51
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:07:56
CYCLE COMPLETE

12/06/2022 09:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____