
Midmark M11 - v1.0.5
Total Cycles: 4499

Name: RITTER2
Sterilizer ID: V2313228
04/09/2025 11:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.6 F 0.0
2:00 120.7 F 0.8
4:00 121.9 F 2.8
6:00 138.7 F 6.1
8:00 167.6 F 10.1
10:00 218.5 F 7.0
12:00 218.7 F 7.0
14:00 226.9 F 11.2
16:00 230.3 F 6.6
18:00 240.9 F 11.9
20:00 238.6 F 10.0
22:00 250.5 F 15.9
24:00 261.3 F 22.0
26:00 271.2 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 271.5 F 29.3
2:00 271.8 F 29.3
3:00 272.5 F 29.5
4:00 272.7 F 29.7

Min 271.0 F 29.3
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.1 F 0.0
5:00 231.7 F 0.0
10:00 223.6 F 0.0
15:00 225.9 F 0.0
20:00 211.9 F 0.0
25:00 169.5 F 0.0
30:00 166.1 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 27:22
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:06:06
CYCLE COMPLETE

04/09/2025 01:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____