
Midmark M11 - v1.0.5
Total Cycles: 2861

Name: RITTER2
Sterilizer ID: V2313228
09/11/2023 01:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.9 F 0.0
2:00 164.4 F 3.1
4:00 188.8 F 7.3
6:00 207.3 F 12.0
8:00 218.8 F 6.1
10:00 232.2 F 10.8
12:00 230.8 F 7.8
14:00 243.5 F 12.5
16:00 244.6 F 12.9
18:00 256.6 F 19.3
20:00 267.3 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.5 F 29.5
2:00 271.5 F 29.0
3:00 272.5 F 29.5
4:00 271.7 F 28.9

Min 271.5 F 28.8
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.9 F 0.0
5:00 208.2 F 0.0
10:00 201.3 F 0.0
15:00 199.6 F 0.0
20:00 195.9 F 0.0
25:00 192.3 F 0.0
30:00 192.1 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 21:27
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:00:45
CYCLE COMPLETE

09/11/2023 02:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____