
Midmark M11 - v1.0.5
Total Cycles: 3735

Name: RITTER2
Sterilizer ID: V2313228
06/26/2024 09:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.2 F 0.0
2:00 146.0 F 1.5
4:00 154.4 F 4.2
6:00 177.2 F 7.9
8:00 196.1 F 11.8
10:00 220.7 F 5.3
12:00 226.1 F 9.6
14:00 236.4 F 10.4
16:00 239.6 F 11.1
18:00 235.5 F 9.1
20:00 249.1 F 15.2
22:00 259.6 F 21.3
24:00 269.8 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.6
1:00 271.8 F 29.5
2:00 272.5 F 29.7
3:00 272.3 F 30.0
4:00 271.8 F 30.1

Min 271.3 F 29.5
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.8 F 0.0
5:00 217.0 F 0.0
10:00 192.9 F 0.0
15:00 191.2 F 0.0
20:00 187.4 F 0.0
25:00 184.9 F 0.0
30:00 184.5 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 25:42
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:04:26
CYCLE COMPLETE

06/26/2024 11:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____