

-----  
Midmark M11 - v1.0.5  
Total Cycles: 873

Name: RITTER2  
Sterilizer ID: V2313228  
03/10/2022 01:25 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 161.3 F 0.0  
2:00 166.6 F 2.8  
4:00 185.5 F 6.8  
6:00 204.2 F 11.5  
8:00 217.9 F 6.0  
10:00 230.1 F 10.8  
12:00 231.9 F 7.8  
14:00 242.9 F 12.4  
16:00 244.5 F 13.0  
18:00 256.4 F 19.4  
20:00 267.2 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.1  
1:00 272.8 F 29.9  
2:00 272.1 F 29.3  
3:00 271.6 F 29.2  
4:00 272.6 F 30.0

Min 271.5 F 29.0  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.9 F 0.0  
5:00 204.4 F 0.0  
10:00 210.4 F 0.0  
15:00 208.3 F 0.0  
20:00 200.3 F 0.0  
25:00 196.8 F 0.0  
30:00 194.0 F 0.0  
Drying Complete

FILLING: 2:12  
HEATING: 21:30  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:01:04  
CYCLE COMPLETE

03/10/2022 02:26 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_