
Midmark M11 - v1.0.0
Total Cycles: 2296

Name: RITTER 1
Sterilizer ID: V2178198
09/07/2022 10:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.9 F 0.0
2:00 173.5 F 1.4
4:00 191.7 F 5.1
6:00 206.0 F 9.5
8:00 224.2 F 7.9
10:00 228.9 F 8.9
12:00 239.2 F 11.8
14:00 239.0 F 10.8
16:00 236.0 F 8.7
18:00 250.7 F 15.4
20:00 262.0 F 21.9
22:00 272.5 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.2
1:00 272.3 F 28.9
2:00 271.5 F 28.2
3:00 271.7 F 28.7
4:00 272.1 F 28.7

Min 271.1 F 28.0
Max 272.3 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.3 F 0.0
5:00 223.9 F 0.0
10:00 231.2 F 0.0
15:00 231.4 F 0.0
20:00 228.1 F 0.0
25:00 228.1 F 0.0
30:00 231.5 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 22:27
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:37
CYCLE COMPLETE

09/07/2022 11:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____