

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4305

Name: RITTER 1  
Sterilizer ID: V2178198  
02/15/2024 11:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.6 F 0.0  
2:00 146.6 F 1.4  
4:00 174.7 F 4.2  
6:00 184.0 F 7.6  
8:00 195.4 F 11.8  
10:00 213.3 F 4.7  
12:00 224.0 F 9.0  
14:00 236.3 F 13.5  
16:00 236.3 F 10.1  
18:00 239.0 F 8.8  
20:00 246.7 F 13.9  
22:00 257.9 F 20.0  
24:00 268.3 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.5 F 29.2  
2:00 272.6 F 29.6  
3:00 271.6 F 28.9  
4:00 272.7 F 29.4

Min 271.5 F 28.9  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.1 F 0.0  
5:00 233.5 F 0.0  
10:00 235.3 F 0.0  
15:00 229.1 F 0.0  
20:00 227.0 F 0.0  
25:00 223.9 F 0.0  
30:00 228.2 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 25:59  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:04:46  
CYCLE COMPLETE

02/15/2024 12:09 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_