

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 003343  
11:57 30/03/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:59  
PRESSURIZING 2:04  
STERILIZING 2:58  
136.0 C 220kPa 2:58  
136.3 C 223kPa 3:28  
136.4 C 226kPa 3:58  
136.4 C 224kPa 4:28  
136.4 C 225kPa 4:58  
136.4 C 225kPa 5:28  
136.4 C 225kPa 5:58  
136.5 C 224kPa 6:28  
136.5 C 224kPa 6:28  
VENTING 6:28  
AIR DRYING 7:21  
CYCLE COMPLETE 37:21  
-----

Digital Signature #  
93122C04463344EA  
-----