
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000314
11:36 27/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:37
PRESSURIZING 3:44
STERILIZING 4:52
136.0 C 227kPa 4:52
136.4 C 230kPa 5:22
136.3 C 232kPa 5:52
136.4 C 233kPa 6:22
136.4 C 233kPa 6:52
136.4 C 233kPa 7:22
136.4 C 230kPa 7:52
136.3 C 232kPa 8:22
136.3 C 232kPa 8:22
VENTING 8:22
AIR DRYING 9:17
CYCLE COMPLETE 18:29

Digital Signature #
B0EAC1543F8525D5
