
Midmark M11 - v1.0.0
Total Cycles: 694

Name: MIDMARK
Sterilizer ID: V2178198
02/23/2021 05:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.4 F 0.0
2:00 132.6 F 1.2
4:00 146.3 F 4.0
6:00 181.5 F 8.2
8:00 222.8 F 11.1
10:00 227.3 F 8.2
12:00 242.4 F 13.5
14:00 240.4 F 10.7
16:00 235.8 F 8.6
18:00 250.8 F 15.6
20:00 263.1 F 22.7
22:00 271.4 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 28.7
1:00 272.5 F 28.8
2:00 272.4 F 28.9
3:00 272.7 F 29.0
4:00 272.6 F 29.2

Min 271.7 F 28.6
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.3 F 0.0
5:00 231.9 F 0.0
10:00 224.3 F 0.0
15:00 226.4 F 0.0
20:00 226.8 F 0.0
Dry Cycle Aborted

FILLING: 1:02
HEATING: 22:36
STERILIZING: 4:00
VENTING: 2:58
DRYING: 22:29
TOTAL CYCLE: 00:53:04

02/23/2021 06:24 PM
ACCEPTED: X
REJECTED:
Operator: U01