
Midmark M11 - v1.0.5
Total Cycles: 4472

Name: RITTER2
Sterilizer ID: V2313228
03/31/2025 09:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.2 F 0.0
2:00 85.8 F 0.2
4:00 80.0 F 1.0
6:00 104.3 F 3.3
8:00 138.6 F 6.3
10:00 173.0 F 9.6
12:00 207.2 F 7.9
14:00 222.3 F 5.4
16:00 230.4 F 8.8
18:00 237.0 F 12.7
20:00 233.3 F 8.7
22:00 243.1 F 13.1
24:00 241.0 F 11.5
26:00 252.3 F 17.2
28:00 262.7 F 23.2
30:00 272.3 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 272.9 F 29.9
1:00 272.0 F 30.0
2:00 272.1 F 30.3
3:00 272.1 F 30.1
4:00 271.5 F 29.9

Min 271.4 F 29.8
Max 272.9 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.3 F 0.0
5:00 229.7 F 0.0
10:00 220.6 F 0.0
15:00 224.9 F 0.0
20:00 225.9 F 0.0
25:00 206.9 F 0.0
30:00 199.8 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 31:01
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:09:25
CYCLE COMPLETE

03/31/2025 10:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____