

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 008345  
9:07 25/02/2025

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:25  
PRESSURIZING 2:27  
STERILIZING 3:26  
136.5 C 222kPa 3:26  
136.4 C 220kPa 3:56  
136.4 C 222kPa 4:26  
136.4 C 220kPa 4:56  
136.2 C 222kPa 5:26  
136.3 C 223kPa 5:56  
136.2 C 220kPa 6:26  
136.4 C 221kPa 6:56  
136.4 C 221kPa 6:56  
VENTING 6:56  
AIR DRYING 7:45  
CYCLE COMPLETE 37:45  
-----

Digital Signature #  
E35CAF9358777B7C  
-----