
Midmark M11 - v1.0.5
Total Cycles: 2058

Name: RITTER2
Sterilizer ID: V2313228
02/15/2023 03:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.1 F 0.0
2:00 198.7 F 3.8
4:00 202.4 F 9.1
6:00 222.1 F 8.5
8:00 233.0 F 11.0
10:00 235.1 F 9.1
12:00 238.5 F 9.6
14:00 249.1 F 15.4
16:00 261.4 F 22.5
18:00 272.6 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.4 F 29.6
2:00 271.5 F 29.3
3:00 272.5 F 29.7
4:00 271.7 F 29.1

Min 271.4 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.4 F 0.0
5:00 227.4 F 0.0
10:00 232.3 F 0.0
15:00 229.5 F 0.0
20:00 227.9 F 0.0
25:00 229.9 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 1:60
HEATING: 18:24
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:37
CYCLE COMPLETE

02/15/2023 04:34 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____