
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000079
12:05 14/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:15
STERILIZING 4:30
136.5 C 225kPa 4:30
136.3 C 224kPa 5:00
136.4 C 225kPa 5:30
136.3 C 222kPa 6:00
136.2 C 223kPa 6:30
136.2 C 223kPa 7:00
136.4 C 222kPa 7:30
136.4 C 226kPa 8:00
136.4 C 226kPa 8:00
VENTING 8:00
AIR DRYING 8:55
CYCLE COMPLETE 23:55

Digital Signature #
48E82629882B1E40
