
Midmark M11 - v1.0.5
Total Cycles: 212

Name: RITTER2
Sterilizer ID: V2313228
07/15/2021 07:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.8 F 0.0
2:00 148.8 F 1.8
4:00 156.8 F 4.7
6:00 181.5 F 8.9
8:00 200.2 F 10.2
10:00 219.0 F 7.8
12:00 231.7 F 12.8
14:00 235.0 F 10.2
16:00 235.6 F 8.8
18:00 249.2 F 15.3
20:00 260.5 F 21.8
22:00 271.1 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.5
1:00 272.2 F 29.4
2:00 272.7 F 29.6
3:00 272.5 F 29.9
4:00 272.0 F 29.8

Min 271.8 F 29.3
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.4 F 0.0
5:00 231.2 F 0.0
10:00 222.6 F 0.0
15:00 225.5 F 0.0
20:00 211.9 F 0.0
25:00 214.0 F 0.0
30:00 212.2 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 23:23
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:01:35
CYCLE COMPLETE

07/15/2021 09:14 AM
ACCEPTED: X
REJECTED:
Operator: U02