
Midmark M11 - v1.0.0
Total Cycles: 821

Name: RITTER 1
Sterilizer ID: V2178198
05/19/2021 09:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.4 F 0.0
2:00 198.7 F 3.0
4:00 206.4 F 8.0
6:00 226.3 F 10.6
8:00 233.4 F 9.5
10:00 235.8 F 8.6
12:00 246.2 F 13.6
14:00 245.0 F 12.4
16:00 257.9 F 19.3
18:00 269.2 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.2 F 28.2
2:00 272.7 F 28.7
3:00 271.8 F 28.3
4:00 272.2 F 28.3

Min 271.6 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.1 F 0.0
5:00 235.3 F 0.0
10:00 237.8 F 0.0
15:00 221.7 F 0.0
20:00 219.2 F 0.0
25:00 222.2 F 0.0
30:00 223.4 F 0.0

Drying Complete

FILLING: 2:15
HEATING: 18:59
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:58:18
CYCLE COMPLETE

05/19/2021 10:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____