
Midmark M11 - v1.0.0
Total Cycles: 1015

Name: RITTER 1
Sterilizer ID: V2178198
07/15/2021 07:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.1 F 0.0
2:00 172.1 F 1.8
4:00 183.2 F 5.3
6:00 217.8 F 10.2
8:00 223.4 F 5.7
10:00 234.5 F 9.8
12:00 236.8 F 9.0
14:00 243.8 F 12.4
16:00 241.1 F 10.7
18:00 253.5 F 16.9
20:00 264.5 F 23.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.8 F 28.8
2:00 271.3 F 28.1
3:00 272.5 F 28.9
4:00 271.9 F 28.3

Min 271.2 F 28.0
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.5 F 0.0
5:00 233.3 F 0.0
10:00 223.8 F 0.0
15:00 227.6 F 0.0
20:00 219.9 F 0.0
25:00 209.9 F 0.0
30:00 206.9 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 21:59
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:31
CYCLE COMPLETE

07/15/2021 08:52 AM
ACCEPTED: X
REJECTED:
Operator: U02