
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002542
8:37 28/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:58
PRESSURIZING 3:02
STERILIZING 4:04
136.1 C 220kPa 4:04
136.4 C 224kPa 4:34
136.0 C 221kPa 5:04
136.4 C 222kPa 5:34
135.9 C 222kPa 6:04
136.4 C 224kPa 6:34
136.2 C 222kPa 7:04
136.2 C 221kPa 7:34
136.2 C 221kPa 7:34
VENTING 7:34
AIR DRYING 8:24
CYCLE COMPLETE 17:10

Digital Signature #
E765A89107FF9505
