
Midmark M11 - v1.0.0
Total Cycles: 2856

Name: RITTER 1
Sterilizer ID: V2178198
02/03/2023 02:44 PM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	80.4 F	0.0
2:00	96.6 F	0.0
4:00	99.4 F	0.1
6:00	138.1 F	0.3
8:00	172.3 F	1.7
10:00	194.9 F	5.6
12:00	208.6 F	10.0
14:00	218.4 F	7.2
16:00	227.9 F	7.9
18:00	239.5 F	12.6
20:00	237.2 F	9.4
22:00	237.9 F	8.8
24:00	248.7 F	14.2
26:00	261.1 F	21.1

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 253F
CHAMBER PRESS 16.4PSI
** ITEMS NOT STERILE **

02/03/2023 03:12 PM