
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003624
9:52 09/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:15
PRESSURIZING 2:16
STERILIZING 3:24
136.4 C 222kPa 3:24
136.0 C 221kPa 3:54
136.4 C 220kPa 4:24
136.4 C 223kPa 4:54
136.4 C 224kPa 5:24
136.3 C 220kPa 5:54
136.3 C 223kPa 6:24
136.4 C 223kPa 6:54
136.4 C 223kPa 6:54
VENTING 6:54
AIR DRYING 7:50
CYCLE COMPLETE 7:53

Digital Signature #
D8998BD6E54832E5
