
Midmark M11 - v1.0.0
Total Cycles: 2457

Name: RITTER 1
Sterilizer ID: V2178198
10/19/2022 11:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.6 F 0.0
2:00 192.8 F 4.5
4:00 205.4 F 9.0
6:00 213.8 F 7.2
8:00 228.2 F 8.4
10:00 242.0 F 13.5
12:00 239.7 F 11.0
14:00 237.5 F 9.3
16:00 251.3 F 16.2
18:00 262.9 F 23.0
20:00 272.1 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.6 F 29.4
2:00 271.7 F 28.6
3:00 272.4 F 29.5
4:00 272.3 F 29.1

Min 271.5 F 28.6
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 230.8 F 0.0
10:00 228.5 F 0.0
15:00 228.2 F 0.0
20:00 228.4 F 0.0
25:00 230.0 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 20:12
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:41
CYCLE COMPLETE

10/19/2022 12:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____