
Midmark M11 - v1.0.0
Total Cycles: 1407

Name: RITTER 1
Sterilizer ID: V2178198
11/30/2021 11:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.9 F 0.0
2:00 150.0 F 2.6
4:00 171.7 F 6.5
6:00 195.3 F 11.4
8:00 216.6 F 4.4
10:00 226.1 F 9.6
12:00 235.0 F 9.0
14:00 240.5 F 12.3
16:00 239.8 F 10.7
18:00 253.3 F 17.5
20:00 265.2 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.8
1:00 272.2 F 29.5
2:00 272.4 F 29.2
3:00 271.5 F 28.8
4:00 272.8 F 29.7

Min 271.4 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.2 F 0.0
5:00 197.1 F 0.0
10:00 213.2 F 0.0
15:00 215.6 F 0.0
20:00 212.1 F 0.0
25:00 209.3 F 0.0
30:00 207.9 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 21:39
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:00:24
CYCLE COMPLETE

11/30/2021 01:04 PM
ACCEPTED: X
REJECTED:
Operator: U02