

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 007375  
6:48 14/05/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:53  
PRESSURIZING 2:54  
STERILIZING 4:02  
136.0 C 220kPa 4:02  
135.7 C 218kPa 4:32  
136.4 C 221kPa 5:02  
136.4 C 219kPa 5:32  
136.1 C 222kPa 6:02  
136.3 C 219kPa 6:32  
136.3 C 221kPa 7:02  
136.4 C 222kPa 7:32  
136.4 C 222kPa 7:32  
VENTING 7:32  
AIR DRYING 8:20  
CYCLE COMPLETE 27:15

-----  
Digital Signature #  
DC232D9563D64193  
-----