
Midmark M11 - v1.0.5
Total Cycles: 2998

Name: RITTER2
Sterilizer ID: V2313228
10/11/2023 01:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.1 F 0.0
2:00 195.5 F 4.3
4:00 209.1 F 10.2
6:00 222.4 F 6.0
8:00 236.9 F 12.2
10:00 238.3 F 10.8
12:00 238.6 F 10.2
14:00 253.6 F 17.7
16:00 266.1 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.0
1:00 272.5 F 29.6
2:00 271.8 F 29.3
3:00 272.2 F 29.3
4:00 272.0 F 29.6

Min 271.5 F 28.9
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 234.5 F 0.0
10:00 221.6 F 0.0
15:00 215.3 F 0.0
20:00 210.1 F 0.0
25:00 220.2 F 0.0
30:00 220.6 F 0.0
Drying Complete

FILLING: 2:37
HEATING: 17:24
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:16
CYCLE COMPLETE

10/11/2023 02:47 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____