

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 000230  
12:17 09/07/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:06  
PRESSURIZING 3:10  
STERILIZING 4:06  
136.0 C 221kPa 4:06  
136.4 C 225kPa 4:36  
136.2 C 223kPa 5:06  
136.2 C 223kPa 5:36  
136.3 C 224kPa 6:06  
136.4 C 225kPa 6:36  
136.2 C 224kPa 7:06  
136.4 C 224kPa 7:36  
136.4 C 224kPa 7:36  
VENTING 7:36  
AIR DRYING 8:26  
CYCLE COMPLETE 8:47

-----  
Digital Signature #  
A6F636B1328E0962  
-----