
Midmark M11 - v1.0.5
Total Cycles: 3151

Name: RITTER2
Sterilizer ID: V2313228
11/21/2023 12:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.6 F 0.0
2:00 179.0 F 3.4
4:00 195.1 F 8.2
6:00 214.9 F 13.3
8:00 223.9 F 7.7
10:00 237.0 F 13.1
12:00 239.4 F 11.2
14:00 239.8 F 10.5
16:00 253.7 F 17.6
18:00 265.8 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.9
1:00 272.9 F 29.3
2:00 272.4 F 29.6
3:00 272.4 F 29.1
4:00 272.6 F 29.7

Min 271.8 F 28.8
Max 273.3 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 206.7 F 0.0
10:00 212.1 F 0.0
15:00 211.8 F 0.0
20:00 208.9 F 0.0
25:00 207.1 F 0.0
30:00 204.5 F 0.0

Drying Complete

FILLING: 1:52
HEATING: 19:36
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

11/21/2023 01:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____