
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000265
16:40 30/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:20
PRESSURIZING 3:24
STERILIZING 4:52
136.4 C 232kPa 4:52
136.2 C 232kPa 5:22
136.3 C 232kPa 5:52
136.3 C 233kPa 6:22
136.4 C 233kPa 6:52
136.4 C 233kPa 7:22
136.4 C 233kPa 7:52
136.4 C 233kPa 8:22
136.4 C 233kPa 8:22
VENTING 8:22
AIR DRYING 9:25
CYCLE COMPLETE 21:20

Digital Signature #
E65F506906ECC953
