

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1924

Name: RITTER 1  
Sterilizer ID: V2178198  
05/10/2022 10:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.3 F 0.0  
2:00 156.1 F 2.0  
4:00 163.9 F 5.3  
6:00 186.5 F 9.7  
8:00 214.1 F 7.0  
10:00 227.9 F 8.8  
12:00 238.4 F 11.3  
14:00 238.4 F 10.5  
16:00 235.8 F 8.0  
18:00 249.3 F 14.8  
20:00 261.1 F 21.4  
22:00 271.7 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.5 F 28.9  
2:00 271.5 F 28.2  
3:00 272.6 F 29.1  
4:00 272.0 F 28.5

Min 271.4 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.0 F 0.0  
5:00 225.0 F 0.0  
10:00 233.7 F 0.0  
15:00 234.6 F 0.0  
20:00 228.3 F 0.0  
25:00 233.8 F 0.0  
30:00 232.5 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 22:36  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:00:37  
CYCLE COMPLETE

05/10/2022 11:41 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_