
Midmark M11 - v1.0.0
Total Cycles: 3741

Name: RITTER 1
Sterilizer ID: V2178198
09/08/2023 11:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.6 F 0.0
2:00 189.5 F 2.5
4:00 206.2 F 7.2
6:00 220.0 F 12.2
8:00 225.5 F 7.1
10:00 237.7 F 12.2
12:00 236.7 F 9.7
14:00 238.6 F 9.3
16:00 248.6 F 14.6
18:00 260.2 F 21.0
20:00 270.6 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.4 F 28.9
2:00 271.5 F 28.5
3:00 272.4 F 28.9
4:00 271.4 F 28.2

Min 271.4 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.3 F 0.0
5:00 222.6 F 0.0
10:00 237.0 F 0.0
15:00 231.8 F 0.0
20:00 222.1 F 0.0
25:00 215.3 F 0.0
30:00 201.0 F 0.0
Drying Complete

FILLING: 2:39
HEATING: 20:47
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:01:02
CYCLE COMPLETE

09/08/2023 12:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____