
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000432
7:23 14/09/2020

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:03
PRESSURIZING 4:05
STERILIZING 5:37
136.0 C 220kPa 5:37
136.4 C 223kPa 6:07
136.3 C 223kPa 6:37
136.2 C 222kPa 7:07
136.4 C 223kPa 7:37
136.3 C 222kPa 8:07
136.3 C 222kPa 8:37
136.4 C 221kPa 9:07
136.4 C 221kPa 9:07
VENTING 9:07
AIR DRYING 10:10
CYCLE COMPLETE 10:16

Digital Signature #
CA41A2CB12B03120
