

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002158  
10:37 28/07/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:46  
PRESSURIZING 2:49  
STERILIZING 3:50  
136.0 C 220kPa 3:50  
136.4 C 222kPa 4:20  
136.4 C 225kPa 4:50  
136.4 C 223kPa 5:20  
136.2 C 222kPa 5:50  
136.4 C 225kPa 6:20  
136.3 C 223kPa 6:50  
136.3 C 224kPa 7:20  
136.3 C 224kPa 7:20  
VENTING 7:20  
AIR DRYING 8:14  
CYCLE COMPLETE 8:17

-----  
Digital Signature #  
D4D7C13FC63225FB  
-----