
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 002273
14:45 15/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:02
STERILIZING 3:00
136.0 C 220kPa 3:00
136.4 C 225kPa 3:30
136.2 C 224kPa 4:00
136.3 C 224kPa 4:30
136.4 C 226kPa 5:00
136.3 C 225kPa 5:30
136.2 C 223kPa 6:00
136.4 C 223kPa 6:30
136.4 C 223kPa 6:30
VENTING 6:30
AIR DRYING 7:22
CYCLE COMPLETE 37:22

Digital Signature #
E19CFBB2F2F5FD95
