
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.1uS / 0.7ppm
CYCLE NUMBER 001086
10:15 20/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:18
PRESSURIZING 3:28
STERILIZING 4:46
136.0 C 221kPa 4:46
136.3 C 224kPa 5:16
136.5 C 225kPa 5:46
136.4 C 225kPa 6:16
136.2 C 223kPa 6:46
136.4 C 224kPa 7:16
136.4 C 223kPa 7:46
136.3 C 223kPa 8:16
136.3 C 223kPa 8:16
VENTING 8:16
AIR DRYING 9:11
CYCLE COMPLETE 24:11

Digital Signature #
7E253646C9CFA2EE
