
Midmark M11 - v1.0.5
Total Cycles: 1320

Name: RITTER2
Sterilizer ID: V2313228
07/19/2022 01:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.1 F 0.0
2:00 143.9 F 0.0
4:00 165.7 F 0.6
6:00 191.1 F 4.6
8:00 211.8 F 9.4
10:00 229.5 F 8.4
12:00 234.6 F 10.3
14:00 235.6 F 8.5
16:00 244.4 F 12.5
18:00 244.4 F 13.0
20:00 256.8 F 19.5
22:00 267.7 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.9
1:00 271.6 F 29.2
2:00 272.7 F 29.9
3:00 272.4 F 29.4
4:00 271.8 F 29.0

Min 271.4 F 28.7
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.2 F 0.0
5:00 213.5 F 0.0
10:00 188.1 F 0.0
15:00 205.4 F 0.0
20:00 220.0 F 0.0
25:00 215.1 F 0.0
30:00 216.6 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 23:18
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:53
CYCLE COMPLETE

07/19/2022 02:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____