
Midmark M11 - v1.0.5
Total Cycles: 795

Name: RITTER2
Sterilizer ID: V2313228
02/14/2022 01:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.8 F 0.0
2:00 167.4 F 3.6
4:00 190.3 F 8.4
6:00 214.5 F 9.1
8:00 227.2 F 9.5
10:00 233.3 F 8.8
12:00 243.2 F 12.5
14:00 245.1 F 13.2
16:00 257.8 F 20.1
18:00 269.7 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.5 F 29.8
2:00 272.3 F 29.3
3:00 271.5 F 28.8
4:00 271.5 F 29.2

Min 271.2 F 28.7
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.6 F 0.0
5:00 216.7 F 0.0
10:00 223.2 F 0.0
15:00 221.5 F 0.0
20:00 222.5 F 0.0
25:00 227.8 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 19:02
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:50
CYCLE COMPLETE

02/14/2022 02:02 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____