

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 006655  
10:39 21/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:38  
PRESSURIZING 2:40  
STERILIZING 3:45  
136.4 C 223kPa 3:45  
136.4 C 224kPa 4:15  
136.4 C 223kPa 4:45  
136.3 C 225kPa 5:15  
136.2 C 221kPa 5:45  
136.1 C 221kPa 6:15  
136.4 C 225kPa 6:45  
136.4 C 226kPa 7:15  
136.4 C 226kPa 7:15  
VENTING 7:15  
AIR DRYING 8:08  
CYCLE COMPLETE 8:17

-----  
Digital Signature #  
5C962FC306528EC4  
-----