
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001415
11:26 14/04/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:53
PRESSURIZING 2:57
STERILIZING 4:06
136.5 C 225kPa 4:06
136.2 C 223kPa 4:36
136.3 C 222kPa 5:06
136.0 C 222kPa 5:36
136.3 C 223kPa 6:06
136.3 C 222kPa 6:36
136.1 C 222kPa 7:06
136.2 C 223kPa 7:36
136.2 C 223kPa 7:36
VENTING 7:36
AIR DRYING 8:23
CYCLE COMPLETE 23:23

Digital Signature #
5DF2E90C3693A4E5
