
Midmark M11 - v1.0.0
Total Cycles: 4857

Name: RITTER 1
Sterilizer ID: V2178198
08/28/2024 08:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.4 F 0.0
2:00 74.7 F 0.0
4:00 81.7 F 0.0
6:00 112.7 F 0.6
8:00 142.2 F 0.7
10:00 173.2 F 2.0
12:00 197.1 F 5.3
14:00 210.5 F 8.6
16:00 219.1 F 12.7
18:00 225.2 F 6.5
20:00 235.8 F 11.0
22:00 232.4 F 7.5
24:00 246.0 F 12.8
26:00 244.4 F 12.3
28:00 256.7 F 18.7
30:00 267.6 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.0 F 28.1
2:00 272.5 F 28.5
3:00 272.3 F 28.7
4:00 271.5 F 27.9

Min 271.4 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 192.7 F 0.0
10:00 190.2 F 0.0
15:00 192.7 F 0.0
20:00 209.5 F 0.0
25:00 216.9 F 0.0
30:00 217.3 F 0.0
Drying Complete

FILLING: 0:41
HEATING: 31:18
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:09:22
CYCLE COMPLETE

08/28/2024 09:38 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____