
Midmark M11 - v1.0.5
Total Cycles: 2426

Name: RITTER2
Sterilizer ID: V2313228
05/29/2023 10:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.1 F 0.0
2:00 191.2 F 3.5
4:00 203.3 F 9.3
6:00 218.9 F 7.4
8:00 233.3 F 11.1
10:00 235.4 F 9.2
12:00 237.7 F 9.3
14:00 249.9 F 15.5
16:00 262.2 F 22.6
18:00 272.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.7
1:00 271.8 F 29.5
2:00 272.5 F 29.5
3:00 272.4 F 29.9
4:00 272.5 F 29.5

Min 271.0 F 28.5
Max 273.4 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.0 F 0.0
5:00 218.1 F 0.0
10:00 239.6 F 0.0
15:00 242.3 F 0.0
20:00 231.7 F 0.0
25:00 223.5 F 0.0
30:00 219.9 F 0.0
Drying Complete

FILLING: 2:19
HEATING: 18:23
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:53
CYCLE COMPLETE

05/29/2023 10:59 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____