
Midmark M11 - v1.0.0
Total Cycles: 2531

Name: RITTER 1
Sterilizer ID: V2178198
11/08/2022 03:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.4 F 0.0
2:00 172.5 F 3.7
4:00 182.9 F 8.1
6:00 204.8 F 13.2
8:00 219.6 F 7.4
10:00 234.6 F 12.7
12:00 234.9 F 9.4
14:00 239.4 F 9.1
16:00 247.3 F 13.9
18:00 259.7 F 20.6
20:00 270.2 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.5 F 29.2
2:00 271.8 F 28.4
3:00 272.1 F 29.1
4:00 272.3 F 29.0

Min 271.4 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.4 F 0.0
5:00 224.8 F 0.0
10:00 231.7 F 0.0
15:00 232.8 F 0.0
20:00 229.0 F 0.0
25:00 229.2 F 0.0
30:00 230.0 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 20:55
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:08
CYCLE COMPLETE

11/08/2022 04:13 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____